Column Replacement Guide



NOTE Column change instructions are only to be used when maintenance is required and are not intended for practice purpose.

INOGENONE

Press the alert/bell button once and screen will display "sieve reset success." Press the power button to turn on the Inogen One[®] G5, and use normally. Press and hold the plus (+) and (-) minus buttons for 5 seconds. The screen will display "sieve reset" message. Release buttons once message is displayed on screen. Push the column assembly into the device such that the columns are fully seated into the Inogen One® G5 concentrator. The spring loaded latch button should fully return to the closed position.

Connect the AC power supply cord to the Inogen One® G5 and plug the power supply AC cord into an electrical outlet. Do Not Power on the Inogen One® G5 concentrator.

CONTINUES ON BACK ->

INOGENONE



If you are using Inogen Connect App, navigate to the Advanced screen, then to Additional Information screen and click on the

American Medical Sales and Rentals 7032 S Revere Parkway, Suite 320 Centennial, CO 80112

$I N O G E N O N E^{\circ G^5}$ Column Change Guide



Phone: 877-774-9271

Column Reset button.

oxygenconcentratorstore.com E-mail: sales@amsrco.com



